



**Jim Doyle, Governor**  
**Mary P. Burke, Secretary**

January 18, 2006

SCICAN  
LUX & ZWINGENBERGER  
BRENDA MURPHY  
1440 DON MILLS RD  
ONTARIO, CANADA,  
M3B 3P9

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: SCICAN  
Product Name: HYDRIM L110W  
Model Number(s): HYDRIM L110W-DO1 and HYDRIM L110W-MO1  
Product File No: 20050911

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of January 2011.

This approval is contingent upon compliance with the following stipulation(s):

- The discharge from this product must discharge to the sanitary drain system through air-gap or air break into a trapped and vented receptor.
- This product has an acceptable built-in air-gap in the water supply to provide protection of the potable water supply.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Michael J. Beckwith, CIPE  
Plumbing Product Reviewer  
phone: 608-266-6742  
fax: 608-267-9566  
e-mail: mbeckwith@commerce.state.wi.us