



April 3, 2009

STRYKER INSTRUMENTS
BECKY DITTY
4100 E MILHAM AVE.
KALAMAZOO MI 49001

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: STRYKER INSTRUMENTS
Product Name: NEPTUNE WASTE MANAGEMENT SYSTEM
Model Number(s): 700-5 DOCKER DETERGENT DISPENSER
Product File No: 20090120

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of April 2014.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.
- The 1-inch diameter drain from this product must directly connect to a properly size, trapped and vent sanitary drain located within a developed length of 10 feet.

This approval supersedes the approval issued on August 29, 2003 under product file number 20030376.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 2030376.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson
Plumbing Product Reviewer
Phone: 608-266-6742
Fax: 608-267-9566
E-mail: jerry.thompson@wi.gov