



June 24, 2010

STERIS CORPORATION  
KENNETH TAYLOR  
2424 W. 23RD ST.  
ERIE PA 16506

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: STERIS CORPORATION  
Product Name: AMSCO CENTURY MEDIUM STEAM STERILZER 26"X26"  
Model Number(s): V-1261, V-1262, V1263, V-1264, P-1261, P-1262, P-1263, SV-1261, SV-1262, SV-1263, SV-1264  
Product File No: 20100241

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of June 2015.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.
- The indirect waste pipe from this product must discharge by means of an air-gap, an air-gap fitting that conforms to ASME A112.1.3 or an air-gap fitting approved by the Department.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson  
Plumbing Product Reviewer  
Phone: 608-266-6742  
Fax: 608-267-9566  
E-mail: jerry.thompson@wi.gov