



SAFETY AND BUILDINGS DIVISION
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay
Scott Walker, Governor
Dave Ross, Secretary

February 7, 2012

ARJO HUNTLEIGH
GARY LOFGREN
50 N GARY AVE
ROSELLE IL 60172

ARJOHUNTLEIGH HOSPITAL EQUIPMENT AB
JOHN WENNBORG
PO BOX 61
ESLOV
SWEDEN

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: ARJOHUNTLEIGH HOSPITAL EQUIPMENT AB
Product Name: BATHING UNIT
Model Number(s): MALIBU
Product File No: 20120069

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of March 2017.

This approval is contingent upon compliance with the following stipulation(s):

- The drain serving this product must have a drainage capacity of at least 16 gallon per minute.
- The water supply serving this product must meet the following conditions: Maximum static pressure of 87 psig, Minimum dynamic pressure of 7.2 psig, Minimum flow per supply pipe of 6.6 gallons per minute, and the temperature of 100 - 176 degrees F. (Recommended 140 degrees F) for the hot water supply.
- The hand-held shower does not require any additional form of backflow protection because it has an air gap discharge.

This approval supersedes the approval issued on March 23, 2007 under product file number 20070067.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 20070067.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson
Plumbing Product Reviewer
Phone: 608-266-6742
Fax: 608-267-9566
E-mail: jerry.thompson@wi.gov