



SAFETY AND BUILDINGS DIVISION  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

June 6, 2012

A-DEC INC.  
ENGINEERING  
TOM LOUISELL  
2601 CRESTVIEW DRIVE  
PO BOX 111  
NEWBERG OR 97132

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: A-DEC INC.  
Product Name: DENTAL ASSISTANTS INSTRUMENT SYSTEM  
Model Number(s): CART-PAC 7004-C  
Product File No: 20120213

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of November 2017.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.

This approval supersedes the approval issued on August 22, 2007 under product file number 20070371.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 20070371.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson  
Plumbing Product Reviewer  
Phone: 608-266-6742  
Fax: 608-267-9566  
E-mail: jerry.thompson@wi.gov