



September 13, 2012

DOTOLO RESEARCH CORP
RAY DOTOLO
PO BOX 14309
CLEARWATER FL 33766-4309

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: DOTOLO RESEARCH CORP
Product Name: COLON HYDROTHERAPY SYSTEM - TOXGEN
Model Number(s): BSC-VV
Product File No: 20120324

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of September 2017.

This approval supersedes the approval issued on September 18, 2007 under product file number 20070433.

This approval is contingent upon compliance with the following stipulation(s):

- The unit must be installed in accordance with the manufacturer's instructions.
- A reduced pressure principle backflow preventer (RPZ) conforming to ASSE 1013 or CAN/CSA B64.4, or an air gap conforming to ASME A112.1.2, must be installed in the fixture supply serving this product. RPZ's shall be installed in accordance to Wisconsin Administrative Code s. SPS 382.41 (5) (f).
- The indirect waste pipe from this product must discharge by means of an air-break or air-gap into a receptor connected to the sanitary plumbing drain system.
- The cross connection control device shall be maintained and tested in accordance to Wisconsin Administrative Code ch. SPS 382.21 (3).

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Engineering Consultant
Safety and Buildings Division
Department of Safety and Professional Services
(608) 267-1401 Phone
(608) 267-9566 Fax
glen.schlueter@wi.gov Email
7:30AM - 4:30PM CT Work Hours

GWS:gws