



SAFETY AND BUILDINGS DIVISION
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

September 12, 2012

ROCHE DIAGNOSTICS
REGIONAL SERVICE
KEVIN JENKINS
9115 HAGUE ROAD
PO BOX 50457
INDIANAPOLIS IN 46250-0457

ROCHE DIAGNOSTICS
ADAM STERLE
9115 HAGUE ROAD
PO BOX 50457
INDIANAPOLIS IN 46250

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: ROCHE DIAGNOSTICS
Product Name: COBAS ANALYZER SERIES
Model Number(s): C6000 AND INTEGRA 400 PLUS
Product File No: 20120329

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of September 2017.

This approval is contingent upon compliance with the following stipulation(s):

- The unit must be installed in accordance with the manufacturer's instructions.
- A reduced pressure principle backflow preventer (RPZ) conforming to ASSE 1013 or CAN/CSA B64.4, or an air gap conforming to ASME A112.1.2, must be installed in the fixture supply serving this product. RPZ's shall be installed in accordance to Wisconsin Administrative Code s. SPS 382.41 (5) (f). The RPZ may be installed upstream of the deionized water source if material compatibility is a concern, however no outlets for intimate human use or consumption shall be located downstream of the RPZ/air gap.
- The indirect waste pipe from this product must discharge by means of an air-break or air-gap into a receptor connected to the sanitary plumbing drain system.
- The cross connection control device shall be maintained and tested in accordance to Wisconsin Administrative Code ch. SPS 382.21 (3).

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Engineering Consultant
Safety and Buildings Division
Department of Safety and Professional Services
(608) 267-1401 Phone
(608) 267-9566 Fax
glen.schlueter@wi.gov Email
7:30AM - 4:30PM CT Work Hours

GWS:gws

SBD-10564-E (N.10/97)

File Ref: 12032901.DOC