



DIVISION OF INDUSTRY SERVICES  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

August 22, 2013

REVISED COPY

FRESENIUS MEDICAL CARE  
PRODUCTS AND HOSPITAL GROUP  
H. MICHAEL OWEN  
2637 SHADELANDS DR  
WALNUT CREEK CA 94598

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: FRESENIUS MEDICAL CARE  
Product Name: FRESENIUS HEMODIALYSIS SYSTEMS  
Model Number(s): 2008T AND 2008K2  
Product File No: 20130267

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of August 2018.

This approval supersedes the approval issued on June 27, 2008 under product file number 20080285.

This approval is contingent upon compliance with the following stipulation(s):

- No external cross connection is required for these devices because they contain an integral air gap that conforms to ASME A112.1.2.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail