



December 11, 2013

A-DEC INC.  
REGULATORY SERVICES  
ROBERT NAPORA  
2601 CRESTVIEW DRIVE  
PO BOX 111  
NEWBERG OR 97132

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: A-DEC INC.  
Product Name: A-DEC 300 SYSTEMS  
Model Number(s): 311, 334, 335, 352, 353 AND 361 WITH SELF- CONTAINED WATER SYSTEM  
Product File No: 20130346

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of December 2018.

This approval supersedes the approval issued on February 5, 2009 under product file number 20080374.

This approval is contingent upon compliance with the following stipulation(s):

- These devices do not require any cross connection control on the potable water supply because the water supply for these devices is a closed, batch filled bottle system.
- Wastewater discharged from these devices must be conveyed to a properly sized receptor via an air gap that conforms to ASME A112.1.2 or ASME A112.1.3.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail