



July 23, 2014

**REVISED COPY**

SCICAN  
LUX & ZWINGENBERGER  
BRENDA MURPHY  
1440 DON MILLS RD  
ONTARIO, CANADA,  
M3B 3P9

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: SCICAN  
Product Name: HYDRIM  
Model Number(s): C61W-DO1, L110W-DO1, L110W-MO1 AND L110W-D01-G4  
Product File No: 20140077

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of July 2019.

This approval is contingent upon compliance with the following stipulation(s):

This approval supersedes the approval issued on December 20, 2010 under product file number 20100490.

This approval is contingent upon compliance with the following stipulation(s):

- The discharge from this product must discharge to the sanitary drain system through air-gap or air break into a trapped and vented receptor.
- This product has an acceptable (i.e. ASME A112.1.2 compliant) built-in air-gap in the water supply to provide protection of the potable water supply.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail