



June 25, 2014

STRYKER INSTRUMENTS
KIM WIERSEMA
4100 E MILHAM AVE.
KALAMAZOO MI 49001

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: STRYKER INSTRUMENTS
Product Name: NEPTUNE 2 ROVER AND DOCKING STATION WASTE MANAGEMENT SYSTEM
Model Number(s): 0702-014-000
Product File No: 20140090

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of June 2019.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap (ASME A112.1.2 or A112.1.3) must be installed on the water supply pipes serving this product.
- Installation of cross connection control shall conform to s. SPS 382.41 (5).
- The maintenance, repair and registration of a reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) shall conform to s. SPS 382.22.
- The drain piping from this product must connect directly to a properly size, trapped and vented sanitary drain located within a developed length of 10 feet.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail