



May 13, 2014

GETINGE GROUP
GETINGE USA INC.
MICHAEL CAIN
1777 E. HENRIETTA ROAD
ROCHESTER NY 14623

GETINGE GROUP
GETINGE SOURCING LLC
ANTHONY POWELL/ LAURA MAJESKI
1777 E. HENRIETTA ROAD
ROCHESTER NY 14623

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: GETINGE GROUP
Product Name: HEALTHCARE VACUUM/GRAVITY STEAM STERILIZERS
Model Number(s): 633HC SERIES
Product File No: 20140112

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of May 2019.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (i.e. RPZ) shall be installed in the hot and cold water supply lines serving these devices. The RPZ shall conform to American Society of Sanitary Engineers (ASSE) Standard 1013.
- The RPZ's shall be installed and maintained as specified in s. SPS 382.41 (4) and s. SPS 382.41 (5) Wis. Adm. Code.
- The RPZ's shall be registered with this department as specified in s. SPS 382.20 (1) (c) and s. SPS 382.20 (13) Wis. Adm. Code.
- The RPZ's shall be tested as specified in s. SPS 382.22 (8) Wis. Adm. Code.
- The drain, waste and vent (i.e. DWV) piping serving these devices must be properly sized to accept the maximum rate of discharge from these devices. The preexisting DWV system must be sufficiently sized to accept the additional wastewater loading generated by these devices.
- The DWV piping serving these devices must be comprised of a material(s) that are stable at the wastewater temperatures generated by these devices. Under normal operating conditions, the wastewater temperature will not exceed 140°F. However, if the cold water supply is shut off, then the wastewater temperature may exceed 200°F. Note, hot water is supplied to these devices only when the integral steam boiler option is used.
- These devices shall discharge wastewater in accordance with s. SPS 382.50 (2) (a) 2.

Getinge Group
May 13, 2014
Page 2 of 2
Product File No.: 20140112

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail