



DIVISION OF INDUSTRY SERVICES  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

May 1, 2014

WATERBIRTH SOLUTIONS INCORPORATED  
MARLA ALTHOUSE  
10644 RHODY DRIVE  
PORT HADLOCK WA 98339

HYDRO SYSTEMS INCORPORATED  
KEVIN STEINHARDT  
29132 AVENUE PAINE  
VALENCIA CA 91355

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: HYDRO SYSTEMS INCORPORATED  
Product Name: ACRYLIC BIRTH POOL  
Model Number(s): AZALEA  
Product File No: 20140115

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of May 2019.

This approval is contingent upon compliance with the following stipulation(s):

- No additional cross connection control is required on the water supply piping serving this device. The reasons are as follows:
  1. The tub faucet conforms to ASME A112.18.1 and provides an air gap that exceeds the requirements of ASME A112.1.2.
  2. The deck mounted hand-held spray faucet conforms to ASME A112.18.1-2012, ASME A112.18.3 and CSA B125.1-12.
- Because this device generates ozone, it shall be installed in a well ventilated area such that the ambient ozone concentration in air does not exceed 0.1 ppm (0.2 mg/m<sup>3</sup>) total weighted average (TWA).

This approval does not in any way endorse or recognize the practice of exposing newborns to waterborne ozone.

Note that ozone may be corrosive to some plumbing components of this device; metals and elastomeric seals in particular.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail