



DIVISION OF INDUSTRY SERVICES  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

October 3, 2014

**REVISED**

GAMBRO RENAL PRODUCTS, INC  
KAE MILLER  
14143 DENVER WEST PARKWAY SUITE 400  
LAKEWOOD CO 80401

GAMBRO DASCO S.P.A.  
KAE MILLER  
VIA MODENESE,30,I-41036 MEDOLLA(MO),  
ITALY

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: GAMBRO DASCO S.P.A.  
Product Name: GAMBRO PHOENIX HEMODIALYSIS SYSTEM  
Model Number(s): 6023006700  
Product File No: 20140176

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of October 2019.

This approval supersedes the approval issued on May 6, 2009 under product file number 20090135.

This approval is contingent upon compliance with the following stipulation(s):

- A cross connection control device acceptable for high hazard shall be installed in the fixture supply serving this product. If required, the cross connection control device shall:
  1. be installed in accordance with s. SPS 382.41 (5) (f) Wis. Adm. Code;
  2. be tested in accordance with s. SPS 382.22 (8); and
  3. be registered with this department in accordance with s. SPS 382.20 (1) (c).

See s. SPS 382.20 Wis. Adm. Code.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail