



July 17, 2014

MAR COR PURIFICATION  
MIKE E. HALLIDAY  
14550 28TH AVENUE N.  
MINNEAPOLIS MN 55447

Re: Description: PFA TUBING AND FITTINGS  
Manufacturer: MAR COR PURIFICATION  
Product Name: MICROFREE PFA TUBING  
Model Number(s): 3/4-INCH AND 1-INCH  
Product File No: 20140212

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an alternate approval to s. SPS 384.30 (4) (e) based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of July 2019.

This alternate approval is contingent upon compliance with the following stipulation(s):

- This tubing is limited to use as water distribution piping in U.S. FDA 510K registered dialysis systems only.
- This tubing shall be installed by MCP certified installers who also are properly licensed Wisconsin plumbers.
- This tubing shall be labeled as follows at intervals not exceeding 12-inches:
  - a. "MCP MicroFree PFA Tubing, 3/4", P/N 3026777", or
  - b. "MCP MicroFree PFA Tubing, 1", P/N 3025076."
- The tubing shall be used within the following pressure and temperature ranges:
  1. < 100 psig @ 77° F (cold water rating)
  2. < 50 psig @ 190° F (hot water rating)

The hot water rating pertains only to dialysis systems that rely on thermal disinfection.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail  
SBD-10564-E (N.10/97)