



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

August 15, 2014

VISTA RESEARCH GROUP, LLC
JIM CHANDLER
1554 TOWNSHIP ROAD 805
PO BOX 321
ASHLAND OH 44805

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: VISTA RESEARCH GROUP, LLC
Product Name: VISTACLEAR DENTAL WATERLINE TREATMENT SYSTEM
Model Number(s): 1000, 1000-23-C, 1000-34-C, 1000-46-C, 1000-47-C, 1000-48-C and 1000-28-HP
Product File No: 20140246

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of August 2019.

This approval supersedes the approval issued on May 3, 2010 under product file number 20090333.

This approval is contingent upon compliance with the following stipulation(s):

- This product is approved as acceptable form of backflow protection of the water supply serving dental units.
- The backflow protection devices in this product must be tested at least once per year. The testing must be done in accordance with the manufacturer's testing requirements as noted in their "Annual Backflow Preventer Valve Testing" procedures. Copy of the test results should be recorded and kept at the installation site. Any repairs or replacement of the backflow preventers should be noted on test report.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail