



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

July 22, 2015

ETHICON INC./ JOHNSON & JOHNSON
ADVANCED STERILIZATION PRODUCTS
BRIAN JOHNSON
13 BLUE JAY LANE
SAINT PAUL MN 55127

ETHICON INC./ JOHNSON & JOHNSON
ADVANCED STERILIZATION PRODUCTS
MICHAEL BLACK
33 TECHNOLOGY DR.
IRVINE CA 92618

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: ETHICON INC./ JOHNSON & JOHNSON
Product Name: EVOTECH ECR SYSTEM (trans id 2575467)
Model Number(s): 50014
Product File No: 20150186

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of July 2020.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air gap (ASME A112.1.2 or ASME A112.1.3) must be installed on the water supply piping serving these devices.
- Installation of the cross connection control device shall conform to s. SPS 382.41 (5) Wis. Adm. Code.
- The maintenance, repair and registration of a reduced pressure principle backflow preventer (ASSE 1013 of CAN/CSA B64.4) shall conform to s. SPS 382.22 Wis. Adm. Code.
- The drain piping from these devices shall connect directly to a properly sized, trapped and vented sanitary drain located within a developed length of 10 ft.
- The installation of these devices shall be performed by a properly licensed Wisconsin plumber.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail