



March 25, 2016

ROCHE DIAGNOSTICS
ADAM STERLE/ MARK SPRUNGER
9115 HAGUE ROAD
PO BOX 50457
INDIANAPOLIS IN 46250

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: ROCHE DIAGNOSTICS
Product Name: (trans id 2680431) COBAS ANALYZER SERIES
Model Number(s): C311, C6000, E411 AND INTEGRA 400 PLUS
Product File No: 20160073

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of March 2021.

This approval supersedes the approval issued on September 12, 2012 under product file number 20120329.

This approval is contingent upon compliance with the following stipulation(s):

- The unit must be installed in accordance with the manufacturer's instructions.
- Cross connection control acceptable for high hazard must be installed in the fixture supply serving these devices. The cross connection control device selected shall be compatible with deionized water.
- Cross connection control devices shall be installed in accordance to s. SPS 382.41 (5) Wis. Adm. Code.
- The indirect waste pipe from this product shall discharge by means of an air-break or air-gap into a receptor connected to the sanitary plumbing drain system.
- The cross connection control device shall be maintained and tested in accordance to s. SPS 382.22 (8) Wis. Adm. Code.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail