



June 14, 2016

ULTRACLEAN SYSTEMS  
BRYAN RUSSELL  
148 DUNBAR AVE., STE A  
OLDSMAR FL 34677

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: ULTRACLEAN SYSTEMS  
Product Name: (trans id 2716141) COUNTERTOP ULTRASONIC CLEANER  
Model Number(s): 1101 COUNTERTOP  
Product File No: 20160159

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of June 2021.

This approval is contingent upon compliance with the following stipulation(s):

- The unit must be installed in accordance with the manufacturer's instructions.
- Cross connection control acceptable for high hazard must be installed in the fixture supplies serving these devices.
- Cross connection control devices shall be installed in accordance to s. SPS 382.41 (5) Wis. Adm. Code.
- The indirect waste pipe from this product shall discharge by means of an air-break or air-gap into a receptor connected to the sanitary plumbing drain system.
- The cross connection control device shall be maintained and tested in accordance to s. SPS 382.22 (8) Wis. Adm. Code, if applicable.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Dept of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail