



DIVISION OF INDUSTRY SERVICES  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

July 29, 2016

GETINGE GROUP  
ARJOHUNTLEIGH  
GORDON ENDICOTT  
2349 W. LAKE STREET  
ADDISON IL 60101

GETINGE GROUP  
ARJOHUNTLEIGH  
JOE SORCI  
2349 W. LAKE STREET  
ADDISON IL 60101

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: GETINGE GROUP  
Product Name: (trans id 2729032) FREEDOM BATH  
Model Number(s): AFyywwsss-US  
(yy = year of manufacture , ww = week of manufacture, sss = serial number)  
Product File No: 20160183

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of July 2021.

This approval supersedes the approval issued on August 31, 2010 under product file number 20100347.

This approval is contingent upon compliance with the following stipulation(s):

- Approval is issued for this product because the design of the product meets the intent of s. SPS 382.41 (3), Wis. Adm. Code that requires adequate protection of the potable water supply. The ASSE 1014 device installed to serve the handheld shower is considered to provide adequate protection of the potable water serving this product.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail