



January 19, 2017

MIELE, INC.
STEVE POLINSKI
9 INDEPENDENCE WAY
PRINCETON NJ 08540

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: MIELE, INC.
Product Name: (trans id 2877071) MIELE DENTAL WASHER-DISINFECTOR AND WASHERS
Model Number(s): PG 8581, PG 8504, PG 8583, PG 8593, PG 8583 DC, PG8535, PG 8536, PG 8056,
PG 8061, G7881, G7883, G7883CD, G7804, G7835, G7836, G7831, G7856 AND G7859
Product File No: 20170007

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of April 2020.

This approval supersedes the approval issued on April 24, 2015 under product file number 20150077.

This approval is contingent upon compliance with the following stipulation(s):

- The discharge from this product must discharge to the sanitary drain system through air-gap or air break into a trapped and vented receptor.
- This product has an acceptable built-in air-gap in the water supply to provide protection of the potable water supply.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail