



SAFETY AND BUILDINGS DIVISION
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

December 28, 2011

ARJO HUNTLEIGH
GARY LOFGREN
50 N GARY AVE
ROSELLE IL 60172

ARJO HOSPITAL EQUIPMENT AB
JOHN WENNBORG
PO BOX 61
S-241 21,ESLOV, SWEDEN

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: ARJO HOSPITAL EQUIPMENT AB
Product Name: PARKER 500
Model Number(s): AL2XXXX-XX
Product File No: 20110384

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of December 2016.

This approval is contingent upon compliance with the following stipulation(s):

- If this fixture is direct connected to the sanitary drain system a flexible drain pipe with a smooth interior shall be used.
- The waste from this product may discharge by means of an air-break or air-gap into a receptor through a drain connected to the sanitary waste system.
- Additional information is included as attachment(s) to this letter; see attachment A.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson
Plumbing Product Reviewer
Phone: 608-266-6742
Fax: 608-267-9566
E-mail: jerry.thompson@wi.gov