



SAFETY AND BUILDINGS DIVISION  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

February 7, 2012

ARJO HUNTLEIGH  
GARY LOFGREN  
50 N GARY AVE  
ROSELLE IL 60172

ARJOHUNTLEIGH HOSPITAL EQUIPMENT AB  
JOHN WENNBORG  
PO BOX 61  
ESLOV  
SWEDEN

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: ARJOHUNTLEIGH HOSPITAL EQUIPMENT AB  
Product Name: MULTICLEAN DISINFECTION SYSTEM  
Model Number(s): BPA 1101-US  
Product File No: 20120068

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of March 2017.

This approval is contingent upon compliance with the following stipulation(s):

- This plumbing product must be installed with the critical level of the vacuum breaker not less than six inches above the highest use of the hose tip of this product.

This approval supersedes the approval issued on March 3, 2007 under product file number 20070066.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 2007.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson  
Plumbing Product Reviewer  
Phone: 608-266-6742  
Fax: 608-267-9566  
E-mail: jerry.thompson@wi.gov