



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

February 15, 2013

STERIS CORPORATION
CHAD BITTNER
5960 HEISLEY ROAD
MENTOR OH 44060

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: STERIS CORPORATION
Product Name: STERIS WASHING SYSTEMS
Model Number(s): VISION CART, RELIANCE CART, GENFOUR SINGLE CHAMBER, SYNERGY
SINGLE CHAMBER, VISION SINGLE CHAMBER AND VISION MULTI-CHAMBER
Product File No: 20130045

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of February 2018.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.
- The indirect waste pipe from this product must discharge by means of an air-gap; an air-gap fitting that conforms to ASME A112.1.3 or an air-gap fitting approved by the Department.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 266-2602 Fax
glen.schlueter@wi.gov E-mail