



DIVISION OF INDUSTRY SERVICES  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

June 20, 2013

GETINGE GROUP  
GETINGE USA INC.  
MICHAEL CAIN  
1777 E. HENRIETTA ROAD  
ROCHESTER NY 14623

GETINGE GROUP  
GETINGE SOURCING LLC  
ANTHONY POWELL/ LAURA MAJESKI  
1777 E. HENRIETTA ROAD  
ROCHESTER NY 14623

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: GETINGE GROUP  
Product Name: VACUUM/GRAVITY STEAM STERILIZERS  
Model Number(s): 833HC SERIES  
Product File No: 20130143

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of June 2018.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer, which conforms to ASSE Standard 1013, shall be installed on the cold water line serving these devices.
- A reduced pressure principle backflow preventer, which conforms to ASSE Standard 1013, shall be installed on the cold water line serving the optional "Water Saver" if used.
- All reduced principle backflow preventers shall be:
  1. installed in accordance with SPS 382.41 (5) Wis. Adm. Code;
  2. registered in accordance with SPS 382.20 (13) Wis. Adm. Code; and
  3. tested in accordance with SPS 382.21 (8) Wis. Adm. Code.
- The wastewater from these devices shall be discharged to a properly sized sanitary drain, waste and vent system and conform to SPS 382.33 (9) (a) 2. and SPS 382.382.10 (1) (f) Wis. Adm. Code.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 266-2602 Fax  
glen.schlueter@wi.gov E-mail