



DIVISION OF INDUSTRY SERVICES  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

July 24, 2013

GETINGE GROUP  
GETINGE USA INCORPORATED  
DON SOCHA  
1777 E. HENRIETTA ROAD  
ROCHESTER NY 14624

GETINGE GROUP  
GETINGE DISINFECTION AB  
JENNY SAHLSTROM-NILSSON  
LJUNGADALSGATAN 11  
PO BOX 1505  
VAXJO SE 35115

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: GETINGE GROUP  
Product Name: 86 SERIES WASHER - DISINFECTOR  
Model Number(s): 8666 AND 8668  
Product File No: 20130171

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of July 2018.

No external cross connection control devices are required to serve these devices. This is because these devices contain an integral air gap that exceeds the requirements of ASME Standard A112.1.2.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail