



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

August 21, 2013

A-DEC INC.
ENGINEERING
TOM LOUISELL
2601 CRESTVIEW DRIVE
PO BOX 111
NEWBERG OR 97132

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: A-DEC INC.
Product Name: PAC 1
Model Number(s): 3420
Product File No: 20130245

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of August 2018.

This approval supersedes the approval issued on December 30, 2008 under product file number 20080373.

This approval is contingent upon compliance with the following stipulation(s):

- This device does not require any external cross connection control because the device is not connected to an external water supply system.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail