



November 5, 2013

METALL ZUG
BELIMED INCORPORATED
JAY UPCHURCH
2325 CHARLESTON REGIONAL PARKWAY
CHARLESTON SC 29492

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: METALL ZUG
Product Name: BELIMED MEDICAL WASHER
Model Number(s): WD 290, WD 250 AND WD 200
Product File No: 20130306

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

- All potable water lines serving these devices must be protected via one of the following:
 1. An air gap that conforms to ASME A112.1.2; or
 2. a reduced pressure principle back flow preventer that conforms to ASSE 1013.
- The drain piping and receptors must be sized to accept the maximum drain flow rate of the device and be comprised of materials suitable for contact with the wastewater from the device. If necessary, the wastewater shall be cooled prior to being discharged such that the temperature of the wastewater is within the working pressure range of the piping conveying the wastewater.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of November 2018.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail