



November 12, 2013

METALL ZUG
BELIMED INCORPORATED
JAY UPCHURCH
2325 CHARLESTON REGIONAL PARKWAY
CHARLESTON SC 29492

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: METALL ZUG
Product Name: BELIMED STEAM STERILIZERS
Model Number(s): MST-V 6-X-X, MST-H GR 9-6-X, MST-V 3-3-6 AND MST-V 5-5-9
Product File No: 20130317

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of November 2018.

This approval is contingent upon compliance with the following stipulation(s):

- All influent water connections to this device shall be protected with high hazard cross connection control. Examples of acceptable cross connection include:
 1. an air gap that conforms to ASME A112.1.2;
 2. a reduced pressure principle backflow preventer (ASSE 1013/CSA B64.1.3).
- The wastewaters discharged by these devices shall be conveyed to a properly sized drain/receptor via an air gap that conforms to ASME A112.1.2.
- The drain piping and receptor shall be comprised of materials suitable for contact with the maximum temperature of the wastewater being conveyed.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail