



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

June 26, 2014

STERIS CORPORATION
RENEE BROWN/Francois Lagace
5960 Heisley Road
Mentor OH 44060

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: STERIS CORPORATION
Product Name: STERIS SINGLE CHAMBER WASHER/DISINFECTOR
Model Number(s): 2000, 3000 AND 5000 SERIES
Product File No: 20140069

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of June 2019.

This approval is contingent upon compliance with the following stipulation(s):

- External cross connection control is not required for these devices because they contain integral air gaps that greatly exceed the requirements of ASME A112.1.2.
- The drain, waste and vent piping serving these products must properly sized to handle the additional wastewater loading and be comprised of materials that are stable when in contact with the maximum wastewater temperature generated by these devices.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail