



May 12, 2015

MEIKO
DWAYNE BECKNELL
1349 HEIL QUAKER BLVD.
LA VERGNE TN 37086

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: MEIKO
Product Name: TOPLINE
Model Number(s): 10, 20, 30 AND 40
Product File No: 20150085

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of May 2020.

This approval is contingent upon compliance with the following stipulation(s):

- These devices do not require external cross connection control.

These devices contain an integral Water Regulation Advisory Scheme (WRAS) "AA" air gaps.

Water is discharged through the air gap into a receptacle which has an unrestricted spillover to atmosphere at all times.

Type AA air gaps are accepted as adequate backflow protection for both backpressure and backsiphonage for category 5 fluids (i.e. fluids which represents a serious health hazard because of the concentration of pathogenic organisms, radioactive or very toxic substances, including any fluid which contains fecal material or other human waste, butchery or other animal waste or pathogens from any other source).

- The drain, waste and vent system serving these devices shall:
 1. Be properly sized to accept the maximum wastewater flow generated by these devices; and
 2. Be constructed of piping materials that are structurally stable at the maximum wastewater temperatures discharged by these devices.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail