



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

April 24, 2015

CANTEL MEDICAL CORPORATION
MEDIVATORS INCORPORATED
MARY P. CEILLE/THOMAS ARCAND
10565 W. PALMERSHEIM DRIVE
HALES CORNERS WI 53130

CANTEL MEDICAL CORPORATION
MEDIVATORS INCORPORATED
MYRON LYSNE
14605 28TH AVENUE NORTH
MINNEAPOLIS MN 55447

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: CANTEL MEDICAL CORPORATION
Product Name: CER OPTIMA SERIES AUTOMATED ENDOSCOPE REPROCESSORS
Model Number(s): CER-1 OPTIMA, CER-2 OPTIMA, DSD201-LT, DSD EDGE AND ADVANTAGE PLUS
Product File No: 20150096

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of June 2017.

This approval supersedes the approval issued on June 13, 2012 under product file number 20120224.

This approval is contingent upon compliance with the following stipulation(s):

- All potable water lines serving these devices shall be protected with a reduced pressure zone (RPZ) cross connection control device that conforms to either American Society of Sanitary Engineering (ASSE) standard 1013 or Canadian Standards Association (CSA) standard B64.1.3.
- The installation of the RPZ shall conform to s. SPS 382.41 (5) (f).
- The RPZ shall be registered and tested in accordance with s. SPS 382.22 (8).
- The discharge from this product must discharge to the sanitary drain system through air-gap or air break into a properly sized, trapped and vented, receptor. The piping material shall be stable under the maximum water temperature discharged by these devices.
- These devices must be installed, maintained and serviced as directed by the manufacturer to perform as advertised.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail

SBD-10564-E (N.10/97)

File Ref: 15009602.DOC