



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

October 15, 2015

ARJOHUNTLEIGH
GARY LOFGREN
50 N GARY AVE SUITE A
ROSELLE IL 60172

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: ARJOHUNTLEIGH
Product Name: RHAPSODY and PRIMO SYSTEM 2000 (trans id 2603889)
Model Number(s): APXXXX1-US and ARXXXX1-US

The "XXXX" designates type of panel (P200, P220 or P300), length of tub shell, and hydrosound or hydromassage.

Product File No: 20150229

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of October 2020.

This approval supersedes the approval issued on January 10, 2011 under product file number 20110012.

This approval is contingent upon compliance with the following stipulation(s):

- No external cross connection control measures are required for these devices. This is because an acceptable level of cross connection control is integral to these devices. These devices meet the intent of the code by providing a method that ensures the vacuum breakers will not be subjected to backpressure due the height of the shower or disinfectant hose.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail