



January 14, 2016

STERIS CORPORATION  
SARAH BROWN  
5960 HEISLEY ROAD  
MENTOR OH 44060-1834

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: STERIS CORPORATION  
Product Name: STERIS CLEANUP SINK (trans id 2646445)  
Model Number(s): CCPS17235, CCPS317235AH, CCPS317735, CCPS317735AH, CCPS319635,  
CCPS319635AH, CCPS319735, CCPS319735AH, CCPS3110835, CCPS3110835AH,  
CCPS3112035, CCPS3112035AH, CCMTS306036 AND CCMTS306036AH  
Product File No: 20150318

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of January 2021.

This approval is contingent upon compliance with the following stipulation(s):

- The laboratory faucet with the serrated nipple shall be protected by a vacuum breaker that conforms to ASSE 1035.
- The water gun line shall be protected by a RPZ that conforms to ASSE 1013.
- Due the prefabricated, adjustable nature of the sinks and distance between sink basins, these sinks (i.e. 1-, 2- or 3-compartment) may be served by a single, properly sized, p-trap. This is based on a previous code interpretation.
- The RPZ's shall be installed in accordance with s. SPS 382.41 (5) Wis. Adm. Code.
- The RPZ's shall be registered, tested, maintained and repaired in accordance with s. SPS 382.22 Wis. Adm. Code.
- The DWV piping down stream of these devices shall be properly sized to handle the additional wastewater flow and of a material suitable for the maximum wastewater temperature generated by these devices.
- These devices shall be installed by a properly licensed Wisconsin plumber.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Steris Corporation  
January 14, 2016  
Page 2 of 2  
Product File No.: 20150318

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail