



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

March 15, 2016

OLYMPUS AMERICA
LEW BECKHAM
3500 CORPORATE PARKWAY
CENTER VALLEY PA 18034

OLYMPUS AMERICA
JASON ASHRAF
350 CORPORATE PARKWAY
CENTER VALLEY PA 18034

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: OLYMPUS AMERICA
Product Name: (trans id 2663805) ENDOSCOPE REPROCESSOR
Model Number(s): OER-PRO
Product File No: 20160028

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of March 2021.

This approval is contingent upon compliance with the following stipulation(s):

- The faucet to which this device is attached must be protected with a hose connection vacuum breaker which conforms to ASSE 1011.

If these devices are used more than 12 hours per day, then high hazard cross connection control acceptable for continuous pressure, backsiphonage and backpressure, shall be used.

- The wastewater from these devices shall be conveyed to the drain, waste and vent (i.e. DWV) system via an air gap. The air gap shall conform to ASME A112.1.2.
- The DWV piping down stream of these devices shall be properly sized to handle the additional wastewater flow and of a material suitable for the maximum wastewater temperature generated by these devices.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail