



April 21, 2016

MEDWATER SYSTEMS
NATE VANZWEDEN
PO BOX 584
KAYSVILLE UT 84037

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: MEDWATER SYSTEMS
Product Name: (trans id 2696754) MEDWATER ANALYZER FEED WATER SYSTEMS
Model Number(s): MW 30 AND MW 70
Product File No: 20160114

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of April 2021.

This approval is contingent upon compliance with the following stipulation(s):

- The drain lines serving the reverse osmosis system and storage tank overflow shall terminate with an air gap that conforms to American Society of Mechanical Engineers (ASME) A112.1.2 or ASME A112.1.3.
- These devices are not approved for residential sale or installation.
- All piping/tubing downstream of the reverse osmosis membrane, on the permeate side, shall be of materials suitable for contact with high purity water.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail